



State of Wisconsin
Governor Tony Evers

Department of Agriculture, Trade and Consumer Protection
Secretary Randy Romanski
Wisconsin Weights and Measures Laboratory

Calibration Certificate

Date Received: March 21, 2025
Date of Calibration: March 24, 2025
Date Issued: March 28, 2025
Date Due: March 24, 2027


Customer: A-1 SCALE COMPANY, INC.
Address: 3287 SHERMAN WAY
SLINGER, WI 53086
Contact: SCOTT KLESPER
Phone: (262) 677-3555

State Test No.: W25-111
Item(s) Submitted: Weight Kit
Manufacturer: Rice Lake
Condition: Good, Acceptable for Calibration
Tolerance Class: NIST HB 105-1 (1990), Class F
Kit Serial #: A1-WS3
Balance ID#: 3,6,7
Procedure Used: NISTIR 6969 (2019), SOP 8
Temperature: 20.4 °C
Relative Humidity: 49.1 %
Pressure: 732.9 mmHg

Nominal Mass	Mass Unit	Serial No.	Conventional Mass Correction (mg)		NIST HB 105-1 (1990), Class F		Uncertainty (mg)	Coverage Factor (k)
			As Found	As Left	As Found	As Left		
5	kg		71	71	Pass	Pass	61	2.08
2	kg		60	60	Pass	Pass	24	2.07
2	kg	*	24	24	Pass	Pass	24	2.07
1	kg		19	19	Pass	Pass	12	2.07
500	g		14.5	14.5	Pass	Pass	8.6	2.06
200	g		15.1	15.1	Pass	Pass	4.8	2.06
200	g	*	14.4	14.4	Pass	Pass	4.8	2.06
100	g		-0.7	-0.7	Pass	Pass	2.4	2.06
50	g		2.9	2.9	Pass	Pass	1.2	2.06
20	g		0.79	0.79	Pass	Pass	0.48	2.06
20	g	*	0.89	0.89	Pass	Pass	0.48	2.06
10	g		0.69	0.69	Pass	Pass	0.24	2.06
5	g		0.07	0.07	Pass	Pass	0.18	2.06
2	g		0.3	0.3	Pass	Pass	0.14	2.06
2	g	*	0.24	0.24	Pass	Pass	0.14	2.06
1	g		0.18	0.18	Pass	Pass	0.11	2.06
500	mg		0.263	0.263	Pass	Pass	0.089	2.09
200	mg		0.202	0.202	Pass	Pass	0.067	2.10
200	mg	*	0.233	0.233	Pass	Pass	0.067	2.10
100	mg		0.005	0.005	Pass	Pass	0.053	2.11
50	mg		0.129	0.129	Pass	Pass	0.043	2.11
20	mg		0.064	0.064	Pass	Pass	0.032	2.11
20	mg	*	0.084	0.084	Pass	Pass	0.032	2.11
10	mg		0.072	0.072	Pass	Pass	0.026	2.10
5	mg		0.051	0.051	Pass	Pass	0.021	2.13
2	mg		0.059	0.059	Pass	Pass	0.015	2.13

The following standard(s) were used: Metric Weight Set WS-2

This report may not be reproduced, except in full, without the written approval of the Lead Metrologist or Laboratory Director.


Paul Masterson, Lead Metrologist


Justin Lien, Laboratory Director



State of Wisconsin
Governor Tony Evers

Department of Agriculture, Trade and Consumer Protection

Secretary Randy Romanski

Wisconsin Weights and Measures Laboratory

Calibration Certificate

Date Received: March 21, 2025
Date of Calibration: March 24, 2025
Date Issued: March 28, 2025
Date Due: March 24, 2027

State Test No.: W25-111
Item(s) Submitted: Weight Kit
Manufacturer: Rice Lake
Condition: Good, Acceptable for Calibration
Tolerance Class: NIST HB 105-1 (1990), Class F
Kit Serial #: A1-WS4
Balance ID#: 3,6,7
Procedure Used: NISTIR 6969 (2019), SOP 8
Temperature: 20.4 °C
Relative Humidity: 49.1 %
Pressure: 732.9 mmHg

Customer: A-1 SCALE COMPANY, INC.
Address: 3287 SHERMAN WAY
SLINGER, WI 53086
Contact: SCOTT KLESPER
Phone: (262) 677-3555

Nominal Mass	Mass Unit	Serial No.	Conventional Mass Correction (mg)		NIST HB 105-1 (1990), Class F		Uncertainty (mg)	Coverage Factor (k)
			As Found	As Left	As Found	As Left		
5	kg		83	83	Pass	Pass	61	2.08
2	kg		80	80	Pass	Pass	24	2.07
2	kg	*	80	80	Pass	Pass	24	2.07
1	kg		7	7	Pass	Pass	12	2.07
500	g		17.5	17.5	Pass	Pass	8.6	2.06
200	g		10.4	10.4	Pass	Pass	4.8	2.06
200	g	*	7.2	7.2	Pass	Pass	4.8	2.06
100	g		8	8	Pass	Pass	2.4	2.06
50	g		1.9	1.9	Pass	Pass	1.2	2.06
20	g		1.62	1.62	Pass	Pass	0.48	2.06
20	g	*	1.57	1.57	Pass	Pass	0.48	2.06
10	g		0.77	0.77	Pass	Pass	0.24	2.06
5	g		0.06	0.06	Pass	Pass	0.18	2.06
2	g		0.43	0.43	Pass	Pass	0.14	2.06
2	g	*	0.88	0.88	Pass	Pass	0.14	2.06
1	g		0.26	0.26	Pass	Pass	0.11	2.06
500	mg		0.279	0.279	Pass	Pass	0.089	2.09
200	mg		0.254	0.254	Pass	Pass	0.067	2.10
200	mg	*	0.266	0.266	Pass	Pass	0.067	2.10
100	mg		0.168	0.168	Pass	Pass	0.053	2.11
50	mg		0.141	0.141	Pass	Pass	0.043	2.11
20	mg		0.113	0.113	Pass	Pass	0.032	2.11

The following standard(s) were used: Metric Weight Set WS-2

This report may not be reproduced, except in full, without the written approval of the Lead Metrologist or Laboratory Director.

Paul Masterson

Paul Masterson, Lead Metrologist

Justin Lien

Justin Lien, Laboratory Director



State of Wisconsin
Governor Tony Evers

Department of Agriculture, Trade and Consumer Protection

Secretary Randy Romanski

Wisconsin Weights and Measures Laboratory

Calibration Certificate

Date Received: March 21, 2025
Date of Calibration: March 24, 2025
Date Issued: March 28, 2025
Date Due: March 24, 2027

Customer: A-1 SCALE COMPANY, INC.
Address: 3287 SHERMAN WAY
SLINGER, WI 53086
Contact: SCOTT KLESPER
Phone: (262) 677-3555

State Test No.: W25-111
Item(s) Submitted: Weight Kit
Manufacturer:
Condition: Good, Acceptable for Calibration
Tolerance Class: NIST HB 105-1 (1990), Class F
Kit Serial #: A1-WS6
Balance ID#: 3,6,7
Procedure Used: NISTIR 6969 (2019), SOP 8
Temperature: 20.4 °C
Relative Humidity: 49.1 %
Pressure: 732.9 mmHg

Nominal Mass	Mass Unit	Serial No.	Conventional Mass Correction (mg)		NIST HB 105-1 (1990), Class F		Uncertainty (mg)	Coverage Factor (k)
			As Found	As Left	As Found	As Left		
5	kg		235	235	Pass	Pass	61	2.08
2	kg		86	86	Pass	Pass	24	2.07
2	kg	*	88	88	Pass	Pass	24	2.07
1	kg		41	41	Pass	Pass	12	2.07
500	g		19.5	19.5	Pass	Pass	8.6	2.06
200	g		13.6	13.6	Pass	Pass	4.8	2.06
200	g	*	16.8	16.8	Pass	Pass	4.8	2.06
100	g		5.5	5.5	Pass	Pass	2.4	2.06
50	g		3.4	3.4	Pass	Pass	1.2	2.06
20	g		1.52	1.52	Pass	Pass	0.48	2.06
20	g	*	1.03	1.03	Pass	Pass	0.48	2.06
10	g		0.99	0.99	Pass	Pass	0.24	2.06
5	g		0.48	0.48	Pass	Pass	0.18	2.06
2	g		0.16	0.16	Pass	Pass	0.14	2.06
2	g	*	0.16	0.16	Pass	Pass	0.14	2.06
1	g		0.22	0.22	Pass	Pass	0.11	2.06
500	mg		0.267	0.267	Pass	Pass	0.089	2.09
200	mg		0.244	0.244	Pass	Pass	0.067	2.10
200	mg	*	0.102	0.102	Pass	Pass	0.067	2.10
100	mg		0.159	0.159	Pass	Pass	0.053	2.11
50	mg		0.07	0.07	Pass	Pass	0.043	2.11
20	mg		0.12	0.12	Pass	Pass	0.032	2.11
20	mg	*	0.072	0.072	Pass	Pass	0.032	2.11
10	mg		0.081	0.081	Pass	Pass	0.026	2.10
5	mg		0.001	0.001	Pass	Pass	0.021	2.13
2	mg		0.006	0.006	Pass	Pass	0.015	2.13
1	mg		0.042	0.042	Pass	Pass	0.013	2.14

The following standard(s) were used: Metric Weight Set WS-2

This report may not be reproduced, except in full, without the written approval of the Lead Metrologist or Laboratory Director.

Paul Masterson, Lead Metrologist

Justin Lien, Laboratory Director



State of Wisconsin
Governor Tony Evers

Department of Agriculture, Trade and Consumer Protection
Secretary Randy Romanski
Wisconsin Weights and Measures Laboratory

Calibration Certificate

Date Received: March 21, 2025
Date of Calibration: March 24, 2025
Date Issued: March 28, 2025
Date Due: March 24, 2027


Customer: A-1 SCALE COMPANY, INC.
Address: 3287 SHERMAN WAY
SLINGER, WI 53086
Contact: SCOTT KLESPER
Phone: (262) 677-3555

State Test No.: W25-111
Item(s) Submitted: Weight Kit
Manufacturer: Rice Lake
Condition: Good, Acceptable for Calibration
Tolerance Class: NIST HB 105-1 (1990), Class F
Kit Serial #: A1-WS12
Balance ID#: 3,6,7
Procedure Used: NISTIR 6969 (2019), SOP 8
Temperature: 20.4 °C
Relative Humidity: 49.1 %
Pressure: 732.9 mmHg

Nominal Mass	Mass Unit	Serial No.	Conventional Mass Correction (mg)		NIST HB 105-1 (1990), Class F		Uncertainty (mg)	Coverage Factor (k)
			As Found	As Left	As Found	As Left		
5	kg		143	143	Pass	Pass	61	2.08
2	kg		55	55	Pass	Pass	24	2.07
2	kg	*	57	57	Pass	Pass	24	2.07
1	kg		29	29	Pass	Pass	12	2.07
500	g		24.5	24.5	Pass	Pass	8.6	2.06
200	g		8.4	8.4	Pass	Pass	4.8	2.06
200	g	*	9.7	9.7	Pass	Pass	4.8	2.06
100	g		7.7	7.7	Pass	Pass	2.4	2.06
50	g		5.6	5.6	Pass	Pass	1.2	2.06
20	g		0.9	0.9	Pass	Pass	0.48	2.06
20	g	*	1.58	1.58	Pass	Pass	0.48	2.06
10	g		0.62	0.62	Pass	Pass	0.24	2.06
5	g		0.58	0.58	Pass	Pass	0.18	2.06
2	g		0.59	0.59	Pass	Pass	0.14	2.06
2	g	*	0.44	0.44	Pass	Pass	0.14	2.06
1	g		0.14	0.14	Pass	Pass	0.11	2.06
500	mg		0.395	0.395	Pass	Pass	0.089	2.09
200	mg		0.27	0.27	Pass	Pass	0.067	2.10
200	mg	*	0.24	0.24	Pass	Pass	0.067	2.10
100	mg		0.197	0.197	Pass	Pass	0.053	2.11
50	mg		0.115	0.115	Pass	Pass	0.043	2.11
20	mg		0.039	0.039	Pass	Pass	0.032	2.11
20	mg	*	0.111	0.111	Pass	Pass	0.032	2.11
10	mg		0.064	0.064	Pass	Pass	0.026	2.10

The following standard(s) were used: Metric Weight Set WS-2

This report may not be reproduced, except in full, without the written approval of the Lead Metrologist or Laboratory Director.


Paul Masterson, Lead Metrologist


Justin Lien, Laboratory Director



State of Wisconsin
Governor Tony Evers

Department of Agriculture, Trade and Consumer Protection
Secretary Randy Romanski
Wisconsin Weights and Measures Laboratory

Calibration Certificate

Date Received: March 21, 2025
Date of Calibration: March 24, 2025
Date Issued: March 28, 2025
Date Due: March 24, 2027


Customer: A-1 SCALE COMPANY, INC.
Address: 3287 SHERMAN WAY
SLINGER, WI 53086
Contact: SCOTT KLESPER
Phone: (262) 677-3555

State Test No.: W25-111
Item(s) Submitted: Weight Kit
Manufacturer: Unknown
Condition: Good, Acceptable for Calibration
Tolerance Class: NIST HB 105-1 (1990), Class F
Kit Serial #: (slotted weights)
Balance ID#: 6&7
Procedure Used: NISTIR 6969 (2019), SOP 8
Temperature: 20.7 °C
Relative Humidity: 50.2 %
Pressure: 733.3 mmHg

Nominal Mass	Mass Unit	Serial No.	Conventional Mass Correction (mg)		NIST HB 105-1 (1990), Class F		Uncertainty (mg)	Coverage Factor (k)
			As Found	As Left	As Found	As Left		
1	kg	29001	39	39	Pass	Pass	12	2.07
500	g	30081	24.5	24.5	Pass	Pass	8.6	2.06
200	g	30079	13.1	13.1	Pass	Pass	4.8	2.06
100	g	30080	-0.5	-0.5	Pass	Pass	2.4	2.06

The following standard(s) were used: Metric Weight Set WS-2

This report may not be reproduced, except in full, without the written approval of the Lead Metrologist or Laboratory Director.


Paul Masterson, Lead Metrologist


Justin Lien, Laboratory Director



State of Wisconsin
Governor Tony Evers

Department of Agriculture, Trade and Consumer Protection

Secretary Randy Romanski

Wisconsin Weights and Measures Laboratory

Calibration Certificate

Date Received: March 21, 2025
Date of Calibration: March 24, 2025
Date Issued: March 28, 2025
Date Due: March 24, 2027

State Test No.: W25-111
Item(s) Submitted: Weight Kit
Manufacturer: Rice Lake
Condition: Good, Acceptable for Calibration
Tolerance Class: NIST HB 105-1 (1990), Class F
Kit Serial #: A1-WS1
Balance ID#: 6&7
Procedure Used: NISTIR 6969 (2019), SOP 8
Temperature: 20.3 °C
Relative Humidity: 47.4 %
Pressure: 733.2 mmHg

Customer: A-1 SCALE COMPANY, INC.
Address: 3287 SHERMAN WAY
SLINGER, WI 53086
Contact: SCOTT KLESPER
Phone: (262) 677-3555

Nominal Mass	Mass Unit	Serial No.	Conventional Mass Correction (mg)		NIST HB 105-1 (1990), Class F		Uncertainty (mg)	Coverage Factor (k)
			As Found	As Left	As Found	As Left		
5	lb	1	10	10	Pass	Pass	28	2.05
5	lb	2	1	1	Pass	Pass	28	2.05
5	lb	3	-3	-3	Pass	Pass	28	2.05
5	lb	4	11	11	Pass	Pass	28	2.05
5	lb	5	14	14	Pass	Pass	28	2.05
1	lb	1	13.5	13.5	Pass	Pass	8.5	2.05
1	lb	2	15.5	15.5	Pass	Pass	8.5	2.05
1	lb	3	12.5	12.5	Pass	Pass	8.5	2.05
1	lb	4	18.5	18.5	Pass	Pass	8.5	2.05
1	lb	5	16.5	16.5	Pass	Pass	8.5	2.05
0.5	lb		7.2	7.2	Pass	Pass	5.6	2.06
0.2	lb		5.6	5.6	Pass	Pass	2.2	2.06
0.2	lb	*	2.3	2.3	Pass	Pass	2.2	2.06
0.1	lb		-0.3	-0.3	Pass	Pass	1.1	2.06
0.05	lb		0.79	0.79	Pass	Pass	0.55	2.06
0.02	lb		-0.42	-0.42	Pass	Pass	0.22	2.06
0.02	lb	*	-0.08	-0.08	Pass	Pass	0.22	2.06
0.01	lb		0.52	0.52	Pass	Pass	0.18	2.06
0.005	lb		0.51	0.51	Pass	Pass	0.15	2.06
0.002	lb		0.07	0.07	Pass	Pass	0.14	2.06
0.002	lb	*	-0.07	-0.07	Pass	Pass	0.14	2.06
0.001	lb		0.285	0.285	Pass	Pass	0.087	2.06

The following standard(s) were used: Avoirdupois Weight Set WS-1

This report may not be reproduced, except in full, without the written approval of the Lead Metrologist or Laboratory Director.

Paul Masterson, Lead Metrologist

Justin Lien, Laboratory Director



State of Wisconsin
Governor Tony Evers

Department of Agriculture, Trade and Consumer Protection
Secretary Randy Romanski

Wisconsin Weights and Measures Laboratory

Calibration Certificate

Date Received: March 21, 2025
Date of Calibration: March 24, 2025
Date Issued: March 28, 2025
Date Due: March 24, 2027

Customer: A-1 SCALE COMPANY, INC.
Address: 3287 SHERMAN WAY
SLINGER, WI 53086
Contact: SCOTT KLESPER
Phone: (262) 677-3555

State Test No.: W25-111
Item(s) Submitted: Weight Kit
Manufacturer: Rice Lake
Condition: Good, Acceptable for Calibration
Tolerance Class: NIST HB 105-1 (1990), Class F
Kit Serial #: A1-WS2
Balance ID#: 6&7
Procedure Used: NISTIR 6969 (2019), SOP 8
Temperature: 20.3 °C
Relative Humidity: 47.4 %
Pressure: 733.2 mmHg

Nominal Mass	Mass Unit	Serial No.	Conventional Mass Correction (mg)		NIST HB 105-1 (1990), Class F		Uncertainty (mg)	Coverage Factor (k)
			As Found	As Left	As Found	As Left		
10	lb		79	79	Pass	Pass	54	2.06
5	lb		54	54	Pass	Pass	28	2.05
2	lb		19	19	Pass	Pass	11	2.06
2	lb	*	21	21	Pass	Pass	11	2.06
1	lb		23.5	23.5	Pass	Pass	8.5	2.05
0.5	lb		17.2	17.2	Pass	Pass	5.6	2.06
0.2	lb		6.4	6.4	Pass	Pass	2.2	2.06
0.2	lb	*	3.7	3.7	Pass	Pass	2.2	2.06
0.1	lb		3.4	3.4	Pass	Pass	1.1	2.06
0.05	lb		1.66	1.66	Pass	Pass	0.55	2.06
0.02	lb		0.15	0.15	Pass	Pass	0.22	2.06
0.02	lb	*	0.72	0.72	Pass	Pass	0.22	2.06
0.01	lb		0.72	0.72	Pass	Pass	0.18	2.06
0.005	lb		0.74	0.74	Pass	Pass	0.15	2.06
0.002	lb		0.15	0.15	Pass	Pass	0.14	2.06
0.002	lb	*	0.39	0.39	Pass	Pass	0.14	2.06
0.001	lb		0.225	0.225	Pass	Pass	0.087	2.06

The following standard(s) were used: Avoirdupois Weight Set WS-1

This report may not be reproduced, except in full, without the written approval of the Lead Metrologist or Laboratory Director.

Paul Masterson, Lead Metrologist

Justin Lien, Laboratory Director



State of Wisconsin
Governor Tony Evers

Department of Agriculture, Trade and Consumer Protection

Secretary Randy Romanski

Wisconsin Weights and Measures Laboratory

Calibration Certificate

Date Received: March 21, 2025
Date of Calibration: March 24, 2025
Date Issued: March 28, 2025
Date Due: March 24, 2027

State Test No.: W25-111
Item(s) Submitted: Weight Kit
Manufacturer: Rice Lake
Condition: Good, Acceptable for Calibration
Tolerance Class: NIST HB 105-1 (1990), Class F
Kit Serial #: A1-WS5
Balance ID#: 6&7
Procedure Used: NISTIR 6969 (2019), SOP 8
Temperature: 20.3 °C
Relative Humidity: 47.4 %
Pressure: 733.2 mmHg

Customer: A-1 SCALE COMPANY, INC.
Address: 3287 SHERMAN WAY
SLINGER, WI 53086
Contact: SCOTT KLESPER
Phone: (262) 677-3555

Nominal Mass	Mass Unit	Serial No.	Conventional Mass Correction (mg)		NIST HB 105-1 (1990), Class F		Uncertainty (mg)	Coverage Factor (k)
			As Found	As Left	As Found	As Left		
10	lb		176	176	Pass	Pass	54	2.06
5	lb		102	102	Pass	Pass	28	2.05
2	lb		27	27	Pass	Pass	11	2.06
2	lb	*	39	39	Pass	Pass	11	2.06
1	lb		21.5	21.5	Pass	Pass	8.5	2.05
0.5	lb		14.2	14.2	Pass	Pass	5.6	2.06
0.2	lb		7.5	7.5	Pass	Pass	2.2	2.06
0.2	lb	*	2.7	2.7	Pass	Pass	2.2	2.06
0.1	lb		3.9	3.9	Pass	Pass	1.1	2.06
0.05	lb		1.39	1.39	Pass	Pass	0.55	2.06
0.02	lb		0.85	0.85	Pass	Pass	0.22	2.06
0.02	lb	*	0.41	0.41	Pass	Pass	0.22	2.06
0.01	lb		0.11	0.11	Pass	Pass	0.18	2.06
0.005	lb		0.41	0.41	Pass	Pass	0.15	2.06
0.002	lb		0.53	0.53	Pass	Pass	0.14	2.06
0.002	lb	*	0.47	0.47	Pass	Pass	0.14	2.06
0.001	lb		0.125	0.125	Pass	Pass	0.087	2.06

The following standard(s) were used: Avoirdupois Weight Set WS-1

This report may not be reproduced, except in full, without the written approval of the Lead Metrologist or Laboratory Director.

Paul Masterson, Lead Metrologist

Justin Lien, Laboratory Director



State of Wisconsin
Governor Tony Evers

Department of Agriculture, Trade and Consumer Protection
Secretary Randy Romanski
Wisconsin Weights and Measures Laboratory

Calibration Certificate

Date Received: March 21, 2025
Date of Calibration: March 24, 2025
Date Issued: March 28, 2025
Date Due: March 24, 2027


State Test No.: W25-111
Item(s) Submitted: Weight Kit
Manufacturer: Rice Lake
Condition: Good, Acceptable for Calibration
Tolerance Class: NIST HB 105-1 (1990), Class F
Kit Serial #: A1-WS11
Balance ID#: 6&7
Procedure Used: NISTIR 6969 (2019), SOP 8
Temperature: 20.3 °C
Relative Humidity: 47.4 %
Pressure: 733.2 mmHg

Customer: A-1 SCALE COMPANY, INC.
Address: 3287 SHERMAN WAY
SLINGER, WI 53086
Contact: SCOTT KLESPER
Phone: (262) 677-3555

Nominal Mass	Mass Unit	Serial No.	Conventional Mass Correction (mg)		NIST HB 105-1 (1990), Class F		Uncertainty (mg)	Coverage Factor (k)
			As Found	As Left	As Found	As Left		
10	lb		189	189	Pass	Pass	54	2.06
5	lb		102	102	Pass	Pass	28	2.05
2	lb		35	35	Pass	Pass	11	2.06
2	lb	*	41	41	Pass	Pass	11	2.06
1	lb		22.5	22.5	Pass	Pass	8.5	2.05
8	oz		21.2	21.2	Pass	Pass	5.6	2.06
0.2	lb		9	9	Pass	Pass	2.2	2.06
0.2	lb	*	9	9	Pass	Pass	2.2	2.06
0.1	lb		2.1	2.1	Pass	Pass	1.1	2.06
0.05	lb		1.56	1.56	Pass	Pass	0.55	2.06
0.02	lb		0.6	0.6	Pass	Pass	0.22	2.06
0.02	lb	*	0.31	0.31	Pass	Pass	0.22	2.06
0.01	lb		0.63	0.63	Pass	Pass	0.18	2.06
0.005	lb		0.5	0.5	Pass	Pass	0.15	2.06
0.002	lb		0.41	0.41	Pass	Pass	0.14	2.06
0.002	lb	*	0.08	0.08	Pass	Pass	0.14	2.06
0.001	lb		0.425	0.425	Pass	Pass	0.087	2.06

The following standard(s) were used: Avoirdupois Weight Set WS-1

This report may not be reproduced, except in full, without the written approval of the Lead Metrologist or Laboratory Director.


Paul Masterson, Lead Metrologist


Justin Lien, Laboratory Director



State of Wisconsin
Governor Tony Evers

Department of Agriculture, Trade and Consumer Protection
Secretary Randy Romanski
Wisconsin Weights and Measures Laboratory

Calibration Certificate

Date Received: March 21, 2025
Date of Calibration: March 25, 2025
Date Issued: March 28, 2025
Date Due: March 25, 2027

State Test No.: W25-111
Item(s) Submitted: Cast Weight
Manufacturer: Rice Lake
Condition: Good, Acceptable for Calibration
Tolerance Class: NIST HB 105-1 (1990), Class F


Customer: A-1 SCALE COMPANY, INC.
Address: 3287 SHERMAN WAY
SLINGER, WI 53086
Contact: SCOTT KLESPER
Phone: (262) 677-3555

Balance ID#: 8
Procedure Used: NISTIR 6969 (2019), SOP 8
Temperature: 20.2 °C
Relative Humidity: 44.2 %
Pressure: 727.4 mmHg

Nominal Mass	Mass Unit	Serial No.	Conventional Mass Correction (mg)		NIST HB 105-1 (1990), Class F		Uncertainty (mg)	Coverage Factor (k)
			As Found	As Left	As Found	As Left		
50	lb	A1-C07	-2,050	110	Fail	Pass	280	2.00
50	lb	A1-C06	-2,010	50	Pass	Pass	280	2.00
50	lb	A1-C04	-1,040	-1,040	Pass	Pass	280	2.00
50	lb	A1-C08	-810	-810	Pass	Pass	280	2.00
50	lb	A1-C03	-1,040	-1,040	Pass	Pass	280	2.00
50	lb	A1-C05	-780	-780	Pass	Pass	280	2.00
50	lb	A1-C09	-2,240	80	Fail	Pass	280	2.00
50	lb	A1LV-C08	-1,700	-1,700	Pass	Pass	280	2.00
50	lb	A1LV-C05	-1,540	-1,540	Pass	Pass	280	2.00
50	lb	A1LV-C03	-770	-770	Pass	Pass	280	2.00
50	lb	A1-C54	470	470	Pass	Pass	280	2.00
50	lb	A1-C53	-420	-420	Pass	Pass	280	2.00
50	lb	A1-C15	-760	-760	Pass	Pass	280	2.00
50	lb	A1-C48	-650	-650	Pass	Pass	280	2.00
50	lb	A1-C58	360	360	Pass	Pass	280	2.00
50	lb	A1-C77	910	910	Pass	Pass	280	2.00
50	lb	A1-C61	900	900	Pass	Pass	280	2.00
50	lb	A1-C13	310	310	Pass	Pass	280	2.00
50	lb	A1-C56	60	60	Pass	Pass	280	2.00
50	lb	A1-C12	50	50	Pass	Pass	280	2.00
50	lb	A1-C14	-140	-140	Pass	Pass	280	2.00
50	lb	A1-C55	940	940	Pass	Pass	280	2.00
50	lb	A1-C71	870	870	Pass	Pass	280	2.00
50	lb	A1-C16	-260	-260	Pass	Pass	280	2.00
50	lb	A1-C11	-1,380	-1,380	Pass	Pass	280	2.00
50	lb	A1-C35	-80	-80	Pass	Pass	280	2.00
50	lb	A1-C59	700	700	Pass	Pass	280	2.00
50	lb	A1-C28	-2,460	100	Fail	Pass	280	2.00
50	lb	A1-C42	-950	-950	Pass	Pass	280	2.00

The following standard(s) were used: 50 lb: W50LB, 25 lb: W25LB, 25000 g: W25KG, 20000 g: W20KG, 10000 g: W10KG

This report may not be reproduced, except in full, without the written approval of the Lead Metrologist or Laboratory Director.


Paul Masterson, Lead Metrologist


Justin Lien, Laboratory Director



State of Wisconsin
Governor Tony Evers

Department of Agriculture, Trade and Consumer Protection
Secretary Randy Romanski

Wisconsin Weights and Measures Laboratory

Calibration Certificate

Date Received: March 21, 2025
Date of Calibration: March 25, 2025
Date Issued: March 28, 2025
Date Due: March 25, 2027

State Test No.: W25-111
Item(s) Submitted: Cast Weight
Manufacturer: Rice Lake
Condition: Good, Acceptable for Calibration
Tolerance Class: NIST HB 105-1 (1990), Class F

Customer: A-1 SCALE COMPANY, INC.
Address: 3287 SHERMAN WAY
SLINGER, WI 53086
Contact: SCOTT KLESPER
Phone: (262) 677-3555

Balance ID#: 8
Procedure Used: NISTIR 6969 (2019), SOP 8
Temperature: 20.2 °C
Relative Humidity: 44.2 %
Pressure: 727.4 mmHg

Nominal Mass	Mass Unit	Serial No.	Conventional Mass Correction (mg)		NIST HB 105-1 (1990), Class F		Uncertainty (mg)	Coverage Factor (k)
			As Found	As Left	As Found	As Left		
50	lb	A1-C46	-3,020	90	Fail	Pass	280	2.00
50	lb	A1LV-C01	-1,120	-1,120	Pass	Pass	280	2.00
50	lb	A1-C45	-1,440	-1,440	Pass	Pass	280	2.00
50	lb	A1-C41	-2,530	70	Fail	Pass	280	2.00
50	lb	A1-C44	-840	-840	Pass	Pass	280	2.00
50	lb	A1-C21	-1,550	-1,550	Pass	Pass	280	2.00
50	lb	A1LV-C02	-930	-930	Pass	Pass	280	2.00
50	lb	A1-C43	-910	-910	Pass	Pass	280	2.00
50	lb	A1-C87	-3,360	80	Fail	Pass	280	2.00
50	lb	A1-C80	-2,210	40	Fail	Pass	280	2.00
50	lb	A1-C84	-2,100	70	Fail	Pass	280	2.00
50	lb	A1-C89	-2,630	60	Fail	Pass	280	2.00
50	lb	A1-C82	-1,080	-1,080	Pass	Pass	280	2.00
50	lb	A1-C85	-3,700	110	Fail	Pass	280	2.00
50	lb	A1-C86	-2,820	30	Fail	Pass	280	2.00
50	lb	A1-C81	-2,630	30	Fail	Pass	280	2.00
50	lb	A1-C83	-3,320	60	Fail	Pass	280	2.00
50	lb	A1-C88	-1,710	-1,710	Pass	Pass	280	2.00
50	lb	A1-C72	-490	-490	Pass	Pass	280	2.00
50	lb	A1LV-C04	680	680	Pass	Pass	280	2.00
50	lb	A1LC-C10	260	260	Pass	Pass	280	2.00
25	lb	A1D2	-470	-470	Pass	Pass	140	2.00
25	lb	21PT	-1,270	70	Fail	Pass	140	2.00
25	lb	A1-D3	440	440	Pass	Pass	140	2.00
25	lb	A1-D4	-360	-360	Pass	Pass	140	2.00
25	kg	A1-MD1	470	470	Pass	Pass	310	2.04
20	kg	A1-ME1	1,350	1,350	Pass	Pass	270	2.02
10	kg	A1-MF1	140	140	Pass	Pass	130	2.02

The following standard(s) were used: 50 lb: W50LB, 25 lb: W25LB, 25000 g: W25KG, 20000 g: W20KG, 10000 g: W10KG

This report may not be reproduced, except in full, without the written approval of the Lead Metrologist or Laboratory Director.

Paul Masterson, Lead Metrologist

Justin Lien, Laboratory Director



State of Wisconsin
Governor Tony Evers

Department of Agriculture, Trade and Consumer Protection
Secretary Randy Romanski
Wisconsin Weights and Measures Laboratory

Calibration Certificate

Date Received: March 21, 2025
Date of Calibration: March 21, 2025
Date Issued: March 27, 2025
Date Due: March 21, 2027

State Test No.: W25-111
Item(s) Submitted: Cast Weight
Manufacturer: Various
Condition: Good, Acceptable for Calibration
Tolerance Class: NIST HB 105-1 (1990), Class F


Customer: A-1 SCALE COMPANY, INC.
Address: 3287 SHERMAN WAY
SLINGER, WI 53086
Contact: SCOTT KLESPER
Phone: (262) 677-3555

Balance ID#: 10
Procedure Used: NISTIR 6969 (2019), SOP 8
Temperature: 20.4 °C
Relative Humidity: 43.8 %
Pressure: 730.0 mmHg

Nominal Mass	Mass Unit	Serial No.	Conventional Mass Correction (mg)		NIST HB 105-1 (1990), Class F		Uncertainty (mg)	Coverage Factor (k)
			As Found	As Left	As Found	As Left		
500	lb	A1-B2	-18,800	-18,800	Pass	Pass	3,000	2.01
500	lb	A1-B1	12,200	12,200	Pass	Pass	3,000	2.01
500	lb	A1-B3	-7,500	-7,500	Pass	Pass	3,000	2.01
500	lb	612-216	-700	-700	Pass	Pass	3,000	2.01

The following standard(s) were used: 500 lb: 90499

This report may not be reproduced, except in full, without the written approval of the Lead Metrologist or Laboratory Director.


Paul Masterson, Lead Metrologist


Justin Lien, Laboratory Director